



# Agilent 6460 Triple Quadrupole LC/MS with Standard ESI Source

## Data Sheet



The Agilent 6460 Triple Quadrupole LC/MS with standard ESI source (option 100) delivers superior sensitivity for trace level analysis. The Agilent 6460 Triple Quadrupole LC/MS achieves sensitivity and resolution specifications with autotune.

Parameter	Measure	Specification
MRM sensitivity Signal-to-Noise ratio (S/N) ESI positive	1 pg of reserpine injected on column, quantifying on the transition $m/z$ 609 to 195	S/N > 10,000:1 Noise 1 $\times$ RMS
MRM sensitivity Signal-to-Noise ratio (S/N) ESI negative	1 pg of chloramphenicol injected on column, quantifying on the transition $m/z$ 321 to 152	S/N > 3,000:1 Noise 1 $\times$ RMS
Mass range		$m/z$ 5 – 3,000
Polarity switching		30 ms
Mass resolution (autotune)	Full width at half maximum	0.7 Da
Mass resolution (manual tune)	Full width at half maximum	0.5 Da
Mass accuracy		0.1 Da from 5 – 1,000 $m/z$ 0.01% from 1,000 – 2,000 $m/z$ 0.02% from 2,000 – 3,000 $m/z$
Mass stability		< 0.1 Da in 24 h
Dynamic range		> $6.0 \times 10^6$
Scan modes		MS scan, MS/MS product ion scan, MRM, MS/MS neutral loss/gain scan and precursor ion scan, SIM
Maximum scan rate		12,500 Da/s
Minimum MRM dwell time		1 ms
MRM transitions		450 per time segment > 13,500 ion transactions per method
Dynamic MRM transitions		4,000 ion transitions per method
Triggered MRM transitions		Up to 10 MRM transitions (primary and secondary) for library search and compound confirmation
Collision cell ion clearance		< 1 ms



**Agilent Technologies**

## General system specifications

Parameter	Specification
Single point of control	Single-point data system method capability with full control of Agilent 1200 Series HPLC systems and 6460C Triple Quadrupole LC/MS System
Time programming	<ul style="list-style-type: none"><li>• Polarity change in time segment</li><li>• Scan and SIM or MRM (plus other modes of data collection)</li><li>• Dynamic and triggered MRM aligns MRMs with compound retention time</li><li>• Solvent divert through calibrant delivery system valve</li></ul>
Wide range of ionization sources	<ul style="list-style-type: none"><li>• Electrospray (ESI)</li><li>• APCI source (Atmospheric Pressure Chemical Ionization)</li><li>• Multimode source (simultaneous ESI and APCI)</li><li>• APPI source (Atmospheric Pressure Photo Ionization)</li></ul>
Autotune	Automated optimization of ion optics and mass axis calibration in positive and negative ion modes using a proprietary tune solution
Solvent declustering	Countercurrent gas
Detector	High-energy conversion dynode and high-gain electron multiplier horn
Vacuum system	Two turbomolecular pumps with one mechanical pump

## Ordering information

### G6460CA Option 100: 6460 Triple Quadrupole LC/MS System with Standard ESI Source

Includes the 6460 Triple Quadrupole Mass Spectrometer, MassHunter Workstation Software with both compliance and method optimization software, a PC, a monitor, and service installation of the system.

The above are not standard installation specifications for the 6460 Triple Quad Option 100. Performance specifications in this document are reviewed for accuracy, but they do not represent the tests and procedures performed at installation, which are described in the Agilent 6400 Series Triple Quad LC/MS System Installation Manual, document G3335-90170 or subsequent version number. See Site Preparation Guide and Service Notes for additional product and specification information.

[www.agilent.com/chem/qqq](http://www.agilent.com/chem/qqq)

Agilent Technologies shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc. 2015-2016  
Published in the U.S.A. April 29, 2016  
5991-5219EN



**Agilent Technologies**